

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/05/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:06
INTERNAL STANDARD	VERIFIED	03:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10

[Handwritten signature]
#820

