

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/13/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:38
INTERNAL STANDARD	VERIFIED	14:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:38
BLANK TEST	0.00	14:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:40
BLANK TEST	0.00	14:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:41
BLANK TEST	0.00	14:42

Signature
#789

