

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/18/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:13
BLANK TEST	0.00	20:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:14
BLANK TEST	0.00	20:15

