

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/31/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:00
INTERNAL STANDARD	VERIFIED	17:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:00
BLANK TEST	0.00	17:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:02
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04

[Handwritten signature]
#820

