

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/22/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:11
INTERNAL STANDARD	VERIFIED	14:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:14
BLANK TEST	0.00	14:15

