

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/30/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:03
INTERNAL STANDARD	VERIFIED	20:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:06
BLANK TEST	0.00	20:07

