

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/19/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:19
INTERNAL STANDARD	VERIFIED	06:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:19
BLANK TEST	0.00	06:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:20
BLANK TEST	0.00	06:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:22
BLANK TEST	0.00	06:22

#820

