

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/28/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:19
INTERNAL STANDARD	VERIFIED	05:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:19
BLANK TEST	0.00	05:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:20
BLANK TEST	0.00	05:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:21
BLANK TEST	0.00	05:22

