

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 10/01/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:15
INTERNAL STANDARD	VERIFIED	23:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:18
BLANK TEST	0.00	23:19

