

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/03/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:06
INTERNAL STANDARD	VERIFIED	12:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:07
BLANK TEST	0.00	12:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:08
BLANK TEST	0.00	12:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:09
BLANK TEST	0.00	12:10

