

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/11/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:52
BLANK TEST	0.00	01:53

[Handwritten signature]
#820

