

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/01/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:31
INTERNAL STANDARD	VERIFIED	10:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:34
BLANK TEST	0.00	10:35

