

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 11/28/2025

OPERATOR NAME:
 MICHAEL A CAMPBELL
 Permit #: 42422
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:55
INTERNAL STANDARD	VERIFIED	05:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:57
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:58
BLANK TEST	0.00	05:59

