

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/07/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:05
INTERNAL STANDARD	VERIFIED	06:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:08
BLANK TEST	0.00	06:09

