

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/20/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:21
INTERNAL STANDARD	VERIFIED	05:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:21
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:24
BLANK TEST	0.00	05:25

