

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 01/02/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:40
INTERNAL STANDARD	VERIFIED	07:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:41
BLANK TEST	0.00	07:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44



John D Bateholts
0714