

SUPERVISOR TEST



WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dnt: 113608

Date: 01/24/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:21
INTERNAL STANDARD	VERIFIED	09:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:22
BLANK TEST	0.00	09:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:23
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	09:24
BLANK TEST	0.00	09:25



P. Bateholts # 718