

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 02/02/2025

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:34
INTERNAL STANDARD	VERIFIED	10:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	10:37
BLANK TEST	0.00	10:38

