

SOLUTION CHANGE

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 03/02/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:27
BLANK TEST	0.00	01:28

*[Handwritten signature]*  
#820

