

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 03/10/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:39
INTERNAL STANDARD	VERIFIED	22:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:40
BLANK TEST	0.00	22:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:41
BLANK TEST	0.00	22:42

*#820*

