

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 03/28/2025

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:32
INTERNAL STANDARD	VERIFIED	12:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:32
BLANK TEST	0.00	12:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35

