

SOLUTION CHANGE

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 04/01/2025

OPERATOR NAME:  
JOHN D BATEHOLTS  
Permit #: 43007  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:46
INTERNAL STANDARD	VERIFIED	08:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:50

*J. D. Bateholts* 0719

