

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 04/04/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:17
INTERNAL STANDARD	VERIFIED	16:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:17
BLANK TEST	0.00	16:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:18
BLANK TEST	0.00	16:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:19
BLANK TEST	0.00	16:20

Michael A Campbell #757

