

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 04/11/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:01
INTERNAL STANDARD	VERIFIED	19:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:02
BLANK TEST	0.00	19:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:03
BLANK TEST	0.00	19:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:04
BLANK TEST	0.00	19:05

*[Handwritten signature]*  
#1820

