

SOLUTION CHANGE

WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113608

Date: 05/06/2025

OPERATOR NAME:  
 MICHAEL A CAMPBELL  
 Permit #: 42422  
 Operator Agency/Dept: WARREN CO SO  
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:05
INTERNAL STANDARD	VERIFIED	06:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08

