

SUPERVISOR TEST

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113608

Date: 05/16/2025

OPERATOR NAME:
 MICHAEL A CAMPBELL
 Permit #: 42422
 Operator Agency/Dept: WARREN CO SO
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:54
INTERNAL STANDARD	VERIFIED	05:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:57
BLANK TEST	0.00	05:58

