

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 05/24/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:35
INTERNAL STANDARD	VERIFIED	18:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:35
BLANK TEST	0.00	18:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:36
BLANK TEST	0.00	18:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:37
BLANK TEST	0.00	18:38

