

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 05/30/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:23
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26

*[Handwritten signature]*  
#820

