

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 06/16/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:18
INTERNAL STANDARD	VERIFIED	19:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:18
BLANK TEST	0.00	19:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:20
BLANK TEST	0.00	19:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:21
BLANK TEST	0.00	19:22

