

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 06/20/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:44
INTERNAL STANDARD	VERIFIED	20:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:47
BLANK TEST	0.00	20:48

Handwritten signature
#870

