

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 06/28/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25510

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:46
INTERNAL STANDARD	VERIFIED	18:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:47
BLANK TEST	0.00	18:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:48
BLANK TEST	0.00	18:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50

Signature
#789

