

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 07/01/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:31
INTERNAL STANDARD	VERIFIED	02:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:32
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:34
BLANK TEST	0.00	02:35

#870

