

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 07/05/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:10
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:12
BLANK TEST	0.00	03:13

Handwritten signature
#1870

