

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 07/18/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:13
INTERNAL STANDARD	VERIFIED	20:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:13
BLANK TEST	0.00	20:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:14
BLANK TEST	0.00	20:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:15
BLANK TEST	0.00	20:16

