

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/02/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:58
INTERNAL STANDARD	VERIFIED	08:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:58
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:01
BLANK TEST	0.00	09:02

[Handwritten signature] 714

