

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 01/10/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:40
INTERNAL STANDARD	VERIFIED	01:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44

*[Handwritten signature]*  
#820

