

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/24/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:52
INTERNAL STANDARD	VERIFIED	09:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:55



John D Bateholts # 714