

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/02/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:55
INTERNAL STANDARD	VERIFIED	22:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:55
BLANK TEST	0.00	22:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:56
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:58
BLANK TEST	0.00	22:59



SGT *[Signature]* #757