

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/07/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:18
INTERNAL STANDARD	VERIFIED	09:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:18
BLANK TEST	0.00	09:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:19
BLANK TEST	0.00	09:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:20
BLANK TEST	0.00	09:21

PT Steven T Bunio # 44983

