

SOLUTION CHANGE

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/01/2025

OPERATOR NAME:
 MICHAEL A CAMPBELL
 Permit #: 42422
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:28
INTERNAL STANDARD	VERIFIED	23:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:30
BLANK TEST	0.00	23:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:32

