

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/10/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:44
INTERNAL STANDARD	VERIFIED	20:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:44
BLANK TEST	0.00	20:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:47
BLANK TEST	0.00	20:48

*Handwritten signature*  
#820

