

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/21/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21



SGT *[Signature]* #757