

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/28/2025

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:45
INTERNAL STANDARD	VERIFIED	11:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:45
BLANK TEST	0.00	11:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:47
BLANK TEST	0.00	11:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:48
BLANK TEST	0.00	11:49

