

SOLUTION CHANGE




WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 04/01/2025

OPERATOR NAME:
 JOHN D BATEHOLTS
 Permit #: 43007
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:22
INTERNAL STANDARD	VERIFIED	09:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:22
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:24
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:25
BLANK TEST	0.00	09:26



P. Bateholts