

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/04/2025

OPERATOR NAME:
J T COON
Permit #: 36043
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:58
INTERNAL STANDARD	VERIFIED	10:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:58
BLANK TEST	0.00	10:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:59
BLANK TEST	0.00	11:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:01
BLANK TEST	0.00	11:02

P.O. J.T. Coon 341

