

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/18/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:20
INTERNAL STANDARD	VERIFIED	18:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:22
BLANK TEST	0.00	18:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:23
BLANK TEST	0.00	18:24

