

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/27/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:39
INTERNAL STANDARD	VERIFIED	01:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:40
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:42

[Handwritten signature]
#820

