

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/06/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:48
INTERNAL STANDARD	VERIFIED	02:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:51
BLANK TEST	0.00	02:52

