

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/24/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:50
INTERNAL STANDARD	VERIFIED	19:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:51
BLANK TEST	0.00	19:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:52
BLANK TEST	0.00	19:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:53
BLANK TEST	0.00	19:54

