

SOLUTION CHANGE

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/01/2025

OPERATOR NAME:  
JEREMY M CRISPELL  
Permit #: 51155  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:13
INTERNAL STANDARD	VERIFIED	11:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:13
BLANK TEST	0.00	11:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:14
BLANK TEST	0.00	11:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:16
BLANK TEST	0.00	11:16

